

Atos

DECLARATION OF CONFORMITY FOR THE PRODUCT(S)

Provox Life Comfort Adhesive

Basic UDI-DI: 7331791-ADH-0-000-0003-CZ

This Declaration of Conformity is issued under the sole responsibility of Atos Medical AB. Atos Medical AB declares that the devices covered by the present declaration are in conformity with the Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices and the requirements specified herein are fulfilled and, clause 6.6 of Schedule 3 to the Australian Therapeutic Goods (Medical Devices) Regulations 2002 including amendments in effect at the issuance date.

Intended use/purpose:

Provox Life Comfort Adhesive is a single use adhesive that provides attachment for Provox Life HMEs and accessories after total laryngectomy.

Hörby, Sweden, date as stated above



.....
Henrik Heringslack, Atos Medical Site Manager
on behalf of the CEO of Atos Medical AB.

Manufacturer: Atos Medical AB
Kraftgatan 8, SE-242 35 Hörby
Sweden

Telephone: +46 (0)415 198 00
Email: Info@atosmedical.com
Web: www.atosmedical.com

SRN number: SE-MF-000000725

Competent Authority Medical Products Agency
Sweden

DECLARATION OF CONFORMITY

FOR THE PRODUCT(S)

7331791-ADH-0-000-0003-CZ

REF	Device name	Class*	GMDN code
8403	Provox Life Comfort Adhesive Oval	I	62175
8094	Provox Life Comfort Experience Oval	I	62175

*Product Classification according to Annex VIII EU Medical Device Regulation 2017/745

In compliance with Therapeutic Goods (Medical Devices) Regulations 2002

Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, the classification rules, and the full quality assurance procedures at each stage, from the design of the device until its final inspection before being supplied.

In compliance with UK Medical Devices Regulations 2002 as amended

Each kind of medical device to which the system has been applied fulfils the applicable obligations imposed by Annex VII and the device meets the provisions of the applicable section of the Medical Devices Regulations.

Authorized representative/UK Responsible Person:

Atos Medical UK Ltd Tottle Road
Cartwright House
Nottingham
Nottinghamshire NG2 1RT
England, United Kingdom

Common Specification(s) as per Article 9, and other Union Legislation(s)

- No relevant Common Specifications as defined in Article 9 to list.
- No relevant Union Legislations to list
- No European Representative

Document Approvals
Approved Date: 2025-12-01

Approval Task Verdict: Approve	HIND.ABUTALEB Hind Palm, Regulatory Affairs Professional (hind.abutaleb-atosmedical@coloplast.com) Issuer 01-Dec-2025 15:03:05 GMT+0000
Approval Task Verdict: Approve	KARNIL Karolina Nilsson, Head of RA, Voice & Respiratory Care (karolina.nilsson-atosmedical@coloplast.com) Regulatory 01-Dec-2025 15:19:58 GMT+0000
Approval Task Verdict: Approve	HENRIK.HERINGSLACK Henrik Heringslack, Site Director (henrik.heringslack-atosmedical@coloplast.com) Management 01-Dec-2025 15:39:10 GMT+0000